

**AMENDMENT NO. 2   APRIL 2023**  
**TO**  
**IS 695 : 2020 ACETIC ACID — SPECIFICATION**  
*(Fourth Revision)*

*(Page 5, Annex A, clause A-3.1.2.2)* — Substitute following for the existing clause:

**‘A-3.1.2.2** The thermometer shall bear a valid calibration certificate from any institution authorised to issue calibration certificate traceable to international or national measurement standards.’